



EC – declaration of conformity



We,

Joh. Stieglmeyer GmbH & CO. KG
Ackerstraße 42
D - 32051 Herford,

hereby declare under sole responsibility as the manufacturer that the product
model named below:

Care bed Soleo; Care bed Libra

in the version submitted complies with the regulations of the EC Directive 93/42/EEC
for Medical Devices, last amended by Directive 2007/47/EC dated 5 September 2007.

It is categorised as a Class 1 active medical device.

The relevant technical documentation is kept by the manufacturer's safety
representative.

To evaluate the conformity to the Directives, all applicable parts of the following
standards were referred to:

Harmonised standards:

EN 14971: 2013-04	Risk Analysis for Medical Products
EN 60601-1: 2006	Safety for Medical Electrical Equipment
EN 60601-1-2: 2007	Electromagnetic Compatibility
DIN EN 60601-1-6: 2010	Medical electrical equipment: Suitability for Intended Use
DIN EN 60601-2-52: 2010	Particular requirements for the safety including essential performance of medical beds

International standards:

IEC 60601-2-52: 2009	Medical electrical equipment: Particular requirements for the basic safety and essential performance of medical beds
----------------------	--

Herford, 2015-12-11

Hans Peter Löw
(Management)

Ralf Wiedemann
(Management)